

LITHIUM ION BATTERY TEST SUMMARY

UN 38.3 Lithium Battery Test Summary for ergoflix® M Battery

1 Battery Manufacturer

Yixing Xinchu Energy Technology Co., Ltd.
No. 82 Xinzhong Road, Xinzhuang Street,
Yixing City, Jiangsu Province, 214200,
China

2**Product Manufacturer**

Nanjing Jin Bai He Medical Apparatur Co., Ltd.
NO.12 Wuge Road, Jiangning District,
Nanjing, Jiangsu, 211113 China
+86 85 78 08 66
info@jhbmedical.com

3 UN38.3 Test Lab

Shanghai Research Institute Of Chemical Industry Testing Center
Shanghai TECH, Chemical Industry Testing Co., Ltd.
Service Hotline: 02152569800
Tel.: +86 2152 8153 77-0713/-0714/-0717/-0718
Fax: +86 2152 8153 77-0716/-0745
West entrance, No. 345 East Yun Ling Road, Shanghai, 200062

4 Test Report Number

RF-SDS235

5 Date of Test Report

2021-08-13

6 Description of Cell or Battery

Polymer lithium-ion battery pack, Model No. ITP 2410
Polymer lithium-ion batteries
24 V / 10000 mAh / 240 Wh, 1800 g – 1950 g
L: 545mm, W: 110mm, T: 38mm

7 UN38.3 Tests Performed and Successfully Passed

T.1, T.2, T.3, T.4, T.5, T.7. and T.8. (Note that T.6 and T.8 are not applicable to batteries.)

8 Assembled Battery Testing Requirements

Not Applicable

9 Edition of UN Manual of Tests and Criteria Used

ST/SG/AC.10/11/Rev.5/Amend. 1&Amend. 2 38.3 UNITED NATIONS
„Recommendations on the Transport of Dangerous GOODS“
Manual of Tests and Criteria ST/SG/AC.10/11/Rev.5/Amend. 1&Amend. 2 38.3

10 Name of Signatory