

LITHIUM ION BATTERY TEST SUMMARY

UN 38.3 Lithium Battery Test Summary for ergoflix® Dura Battery

1 Battery Manufacturer

Yixing Xinchu Energy Technology Co., Ltd.
No.82 Xinzhong Road, Xinzhuang Street,
Yixing City, Jiangsu Province, 214200,
China

2 Product Manufacturer

Anhui JBH Medical Apparatus Co., Ltd.
No. 116 Qicang Road, Mingguang, Chuzhou
Anhui Province, Anhui, 239400
CHINA
+86 85 78 08 66
info@jbhmedical.com

3 UN38.3 Test Lab

Shanghai Research Institute Of Chemical Industry Testing Center
Shanghai TECH, Chemical Industry Testing Co., Ltd.
Service Hotline: 02152569800
Tel.: +86 2152 8153 77-0713/-0714/-0717/-0718
Fax: +86 2152 8153 77-0716/-0745
West entrance, No. 345 East Yun Ling Road, Shanghai, 200062

4 Test Report Number

No. 1119040349

5 Date of Test Report

2019-04-25

6 Description of Cell or Battery

High Power Battery Pack D10A/D15
Lithium-ion battery
24V/ 6000mAh / 144Wh, 1484g
L: 230mm, W: 155mm, H: 47mm

7 UN38.3 Tests Performed and Successfully Passed

T.1, T.2, T.3, T.4, T.5, T.7. and T.8. (Note that T.6 and T.8 are not applicable to batteries.)

8 Assembled Battery Testing Requirements

Not Applicable

9 Edition of UN Manual of Tests and Criteria Used

ST/SG/AC.10/11/Rev. 6 Section 38.3 UNITED NATIONS
„Recommendations on the Transport of Dangerous GOODS“
Manual of Tests and Criteria ST/SG/AC.10/11/Rev. 6 Section 38.3

10 Name of Signatory

